





## **GUIDELINES FOR PARTICIPATION IN THE CALIBRATION PROJECT**

## **Description of the Project:**

This project is implemented as part of the Project "Metrology for Sustainable Energy Technologies and the Environment (M4SET)" executed by the Organization of American States (OAS) and the National Institute of Standards and Technology of the United States (NIST).

The objective of this project is to give better traceability to the Ozone measurements in Latin American and Caribbean (LAC) Cities. The project contemplates a series of activities that will be implemented in a span of three months beginning in August and ending in October 2020. These activities include three webinars that will guide the project implementation process and address training needs, an inter-comparison using a level 2 transfer standard calibrated with NIST's SRP primary standard and subsequently calibrated against the Laboratory SRP primary standard of air quality (CALAIRE) in Colombia and the National Institute of Ecology and Climate Change (INECC) in Mexico, and finally, the calibration of photometers or calibrators of the participating cities (beneficiaries of the project) against the primary SRP standard of Cities of Medellin (CALAIRE) and Mexico (INECC).

## **Guidelines for the application:**

The institutions that are interested in participating in the calibration project of a level 2 transfer standard must demonstrate that they comply with the requirements established in the application form.

The number of selected cities is subject to the availability of financial resources. The OAS will cover the calibration costs and, in some cases, may cover the transportation costs of the equipment to the cities where the calibration will be carried out (Medellin or Mexico City). However, please note that those institutes that are willing to cover the cost of transporting the equipment will have priority in the selection. Institutions interested in participating in this project covering their own calibration costs at CALAIRE Air Quality Laboratory will enjoy the same discount granted to the project for the calibration of the UV photometer.

Although several institutions within the same country can participate and benefit from the project, the selection of the institutions that will receive financial support with the payment of the calibration will be made seeking to benefit the greatest number of countries.

The results of the project will be compiled in a publication that will address the details of the calibration and will serve as an-input to identify opportunities for improvement of the monitoring networks. In this sense, the chosen cities must authorize the use of the data resulting from the Project for the publication.







The application deadline is August 20, 2020. The OAS, NIST, and Certifying Bodies will review the proposals and select the beneficiary institutions based on the following selection criteria:

- a. Compliance with the requirements established for the equipment.
- b. Ability to cover the transportation costs of the equipment, or if you cannot meet this cost, state it.
- c. Authorization of the use of the results for the development of the publication.
- d. Number of applications per country.

If you have any questions about the application or the objectives of the project, please contact Bibiana Serna (<a href="mailto:BSerna@oas.org">BSerna@oas.org</a> ), Magdalena Navarro (<a href="mailto:magdalena.navarro@nist.gov">magdalena.navarro@nist.gov</a> ) via email.

Find below the information that must be filled out by the applicants

## M4SET Application Form

Once you complete your form, please send the application to Bibiana Serna (<u>Bserna@oas.org</u>), Magdalena Navarro (<u>mnavarro@nist.gov</u>). Include the phrase "M4SET project" in the subject line of the email.

Application date	
Institution	
Address	
City	
Country	
Responsible for the	
application:	Name:
	Telephone:
	Email:
	Information about the equipment







Bran	d of the			
equip	oment			
Mode	el			
Seria	l Number			
Acqu	isition year			
A 1 1:				
	tional equipment ons installed by			
	nanufacturer			
	tional equipment			
	mation that			
cons	ider important			
	Mark "yes" or "n	o", to indicate if	the equip	oment complies with the indicated characteristic
Item	Description of the ena		Yes/No	Comments / Clarification
1	Does the equipme			
		photometer with a UV		
2	wavelength of 254 nm?			
2	D (1 :			
	Does the equipme	ent have an		
	Does the equipme internal ozone ge	ent have an		
3	internal ozone ge	ent have an nerator? *		
3	Are the operating the equipment wi	ent have an nerator? *  parameters of ithin the values		
	Are the operating the equipment wistipulated by the	ent have an nerator? *  parameters of thin the values manufacturer?		
3	Are the operating the equipment wistipulated by the Is the UV lamp pe	ent have an nerator? *  parameters of ithin the values manufacturer?  rformance at		
	Are the operating the equipment wistipulated by the Is the UV lamp peleast 15% above to	ent have an nerator? *  parameters of thin the values manufacturer?  rformance at the		
	Are the operating the equipment wistipulated by the Is the UV lamp per least 15% above to manufacturer's re-	ent have an nerator? *  parameters of thin the values manufacturer?  rformance at the		
4	Are the operating the equipment wistipulated by the Is the UV lamp peleast 15% above manufacturer's reminimum value?	ent have an nerator? *  parameters of ithin the values manufacturer?  rformance at the ecommended		
	Are the operating the equipment wistipulated by the Is the UV lamp per least 15% above manufacturer's reminimum value?	ent have an nerator? *  s parameters of thin the values manufacturer?  rformance at the ecommended		
4	Are the operating the equipment wistipulated by the Is the UV lamp peleast 15% above manufacturer's reminimum value?	ent have an nerator? *  s parameters of thin the values manufacturer?  rformance at the ecommended		







6	Can you perform internal		
	cleaning of the equipment		
	maintaining the manufacturer's		
	operating parameters before		
	shipping it?		
7	Can you perform external		
	cleaning of the equipment befor	e	
	shipping it?		
8	Is the computer chassis in good		
	condition?		
9	Are the pneumatic connections		
	of the equipment in good		
	condition?		
4.0			
10	Since purchasing the equipment	·	
	have you performed calibration	S	
	with a primary SRP standard?		
	Note: In the observations field		
	indicate how many calibrations		
	and the date of the last one		
	carried out.		
*No+	o. Itom 2 is an assential factor for	the equipp	cont to be calibrated by the Mayigan CDD at the INECC
			nent to be calibrated by the Mexican SRP at the INECC,
	ne Colombian SRP it is not an ena		
Pieas	se answer the following questions	related to	criteria b, c
11	¿Can your institution	Yes	
11	dispense with the photometer	165	
	or calibrator for the time it	No	
	takes for transfers and	110	
	certification in CALAIRE	Observatio	on / Clarification:
	(Medellin) or INNEC (Mexico	Observation	on / Garmeation.
	City)?		
12	What calibration laboratory	Medelli	n – Colombia (CALAIRE)
12	would the equipment be		ii dololiida (drimiia)
	shipped? In your choice,	México	- Mexico City (INEC)
	please note that the	1-10M100	
	calibration location should be		
	selected based on the minimum cost of		







13	Can your institution cover transportation costs of the equipment, to the calibral laboratory in Colombia of Mexico?	ne ition	Yes No Obser	
14	Can your institution cover equipment calibration countries (The benefit would be to participate in the publication of the comparative result the member countries of project)	ests? Ition ats of	Yes No	<u> </u>
15	How many weeks of preparation from August does it take to ship the equipment?	:1		
16	<u> </u>		Yes No Name Positi Institu	on:
Addit Item	Equipment	ne equi		Equipment description Informative only (not enabling)  Comments / Clarification
17	Description  Is the equipment a dynamic Calibrator?  Is the equipment a	,		·
10	photometric ozone calibrator? Does it calibrate only ozone? Such as the Teledyne T703 model.			







19	Does the equipment have the configuration option about the correction of temperature and pressure-				
	activated? Is this option offered by				
	the manufacturer?				
20	If the temperature and pressure option is				
	deactivated, does it allow making the				
	change before				
	performing the				
	calibration?				
21	Have you				
	performed leak tests or ozone leak				
	tests on your				
	equipment?				
22	Do you need advice				
	to carry out the				
	ozone leak / leak				
	test on your equipment?				
23	Can you share the				
	sheet of the last				
	calibration of your				
	equipment? In case				
	it is available.				
Document	completed by:		Signature:		
Position:					
Institute /	organization				